CLEANING and DISINFECTING AIREZ

Emptying and cleaning AIREZ products is easy. The below is a quick reference for both cleaning and disinfecting AIREZ 2.0 cushions, back supports, full body systems, and custom foot boxes.

BEST PRACTICE FOR CLEANING



- Empty cushion with cover still on
- Remove the cover
- Place the cushion or back in a sink or tub with warm, soapy water and a mild detergent
- Swish and soak for at least 5 minutes
- Rinse thoroughly with clean water using a handheld shower or sink sprayer
- Gently shake off excess water and set the product level and upright on an absorbent surface to drain
- For faster drying, place in front of a fan
- Once dry, replace the cover

BEST PRACTICE FOR DISINFECTING

- Remove the cover and liner
- Remove any visible debris or contamination
- Thoroughly spray the cushion, collection tray, back support, and exposed hardware with MadaCide-1
- Allow all surfaces to remain visibly wet for 10 minutes

  NOTE: A common mistake is spraying a disinfectant and immediately wiping it away. To achieve hospital-grade disinfection, MadaCide-1 must remain visibly wet on the surface for the recommended contact time of approximately 0 minutes.
- Wipe dry with a clean cloth
- Reassemble once all components are completely dry

COVER CARE

- Machine wash gentle, cold, with mild detergent
- Do not bleach
- Hang to dry

WHAT TO AVOID

Chemical solvents dissolve or break down other substances. In cleaning products, solvents are often used to remove grease, oil, or adhesives, but they can also damage materials like elastomers, foams, or fabrics.

Common chemical solvents include:

- Isopropyl alcohol
- Acetone
- Bleach
- Turpentine
- Mineral spirits
- Ammonia-based cleaners
- Hydrogen peroxide

These chemicals can cause discoloration, hardening, or degradation of cushion materials and covers, which is why we recommend using only mild soap and water for cleaning and MadaCide-1 for disinfecting.

matrix seating usa

www.matrixseatingusa.com | ph. 800.986.9319 | info@matrixseatingusa.com